



Quality Control Protocol on SPECT

**Ahmad Bitarafan Rajabi,
Professor of medical physics**

Rajaie Cardiovascular Institute

Cardiovascular Intervention Research Center

Cardiac Electrophysiology Research Center

Echocardiography Research Center

Department of nuclear medicine and molecular imaging.

Factors affecting image formation

- **Distribution of radiopharmaceutical**
- **Collimator selection and sensitivity**
- **Spatial resolution, Distance collimator-patient**
- **Energy resolution**
- **Uniformity**
- **Count rate performance**
- **Spatial positioning at different energies**
- **Center of rotation**
- **Scattered radiation**
- **Attenuation**
- **Recording system**
- **Type of examination**
- **Noise, Background**

Image quality depends on:

- **Radiopharmaceutical**

- Labeling
- Administered Dose

- **Technical factors**

- Equipment used
- Acquisition protocol
- Image processing
- Noise
- Spatial resolution
- Scatter

- **Patient factors**

- Size , Age, Disease , Movement



Quality Control in 3 steps

Measurement:

calculation of values related to system

Assessment:

qualitative evaluation to say is it acceptable or not

Correction/Calibration:

modifying parameters to reduce systematic errors

Acceptance Test vs. QC

Acceptance test:

After installation, and before clinical use
to **verify that instrument** performs according to its specification

Each instrument is supplied with a set of basic specifications produced by the manufacturer according to standard test procedures (e.g. NEMA)

Quality control:


Once the instrument has **been accepted for clinical use**

Procedures that are sensitive to changes in performance

Tests must be performed by appropriately qualified and trained staff

- Detailed local operating procedures should be Written for this routine work
- All test results must be **recorded and monitored for variation**
- Appropriate actions taken when changes are observed

Quality Control – Gamma Cameras



	Acceptance	Daily	Weekly	Yearly
Uniformity	P	T	T	P
Uniformity, tomography		P		P
Spectrum display	P	T	T	P
Energy resolution	P			P
Sensitivity	P		T	P
Pixel size	P		T	P
Center of rotation	P		T	P
Linearity	P			P
Resolution	P			P
Count losses	P			P
Multiple window pos	P			P
Total performance phantom	P			P

P: physicist
T: technician

IAEA-TRS-454
Quality
Assurance for
Radioactivity
Measurement
in Nuclear
Medicine 2006

IAEA-TECDOC-602

*Quality control of
Nuclear medicine instruments 1991*



INTERNATIONAL ATOMIC ENERGY AGENCY

IAEA

May 1991


IAEA QA
for SPECT
systems
(in press)

QC Gamma camera

NEMA

**Standards for
Performance Measurement
of Scintillation Cameras...**

**And What They Can Mean
For You**



AAPM REPORT No. 6

**SCINTILLATION CAMERA
ACCEPTANCE TESTING AND
PERFORMANCE EVALUATION**



Published for the American Association of Physicists in Medicine
by the American Institute of Physics

The Hospital Physicists' Association


**The Theory, Specification and
Testing of
Anger Type Gamma Cameras**

Topic Group Report — 27



National Electrical Manufacturers Association

SOURCES FOR QC OF GAMMA CAMERAS

- **Point source**
- **Collimated line source**
- **Line source**
- **Flood source**



<1 mm



Tc99m, Co57, Ga67

QC Activities/ Settings

Intrinsic tests usually 35 – 500 micro Ci.

Extrinsic tests usually 2 – 10 mCi.




Energy window is usually 15 % of Photopeak, if nothing recommended by vendor.

Count rate < **30.000** cps



Phantoms for QC of gamma cameras

- Bar phantom
- Slit phantom
- Orthogonal hole phantom
- Total performance phantom



Phantoms for QC of gamma cameras






Figure A.5.a. 90° Bar quadrant phantom.



Intrinsic/Extrinsic Uniformity



For a uniform flux we shall expect a uniform image!

Unfortunately not!

The most sensitive parameter, Why?

Photopeak Window

Photomultiplier tube (PMT) performance

Imperfections in the collimators

Variations in crystal response


Variations in PMTs' response

Minor fluctuations in the electrical circuitry


Intrinsic/Extrinsic Uniformity



Week 1



Week 2



Week 3



Week 4

Non uniformity due to a drift in circuitry

The single **most important** QC test
should be performed **Daily!**

Intrinsic Uniformity

The uniformity of the camera's response can be checked by imaging a **Tc-99m flood source**.


A **5-30 million count** image of the flood is collected.

A **300-500 micro_Ci** point source.

Count rate < **30000 cps**.

64 x 64 Matrix size.

* A minimum of **10,000** counts shall be collected in the center pixel of the image



Extrinsic Uniformity Visual

The uniformity of the camera's response can be checked by imaging a **Tc-99m** flood source.



A **5-30 million count** image of the flood is collected with **LEHR** Collimator.

A **5-10 mCi** point source.

Count rate < **30000 cps**, **512 x 512** Matrix size.

Uniformity

- Field uniformity may be done either as intrinsic or extrinsic
- **Intrinsic**=collimators off. Monitors the condition of sodium iodide crystals and electronics.
- **Extrinsic**=collimators on. Monitors the camera as it is used clinically.

Uniformity

Uniformity images must be inspected **daily** for nonuniformity and **compared to** previous flood (uniformity) images.


UNIFORMITY

Flood source (Tc-99m, Co-57)

Point source (Tc-99m)

- **Intrinsic uniformity:** Point source at a large distance from the detector. Acquire an image of 10.000.000 counts
- **With collimator:** Flood source on the collimator. Acquire an image of 10.000.000 counts

Uniformity



1. Subjective evaluation of the image
2. Calculate
 - Integral uniformity (IU)
 - Differential uniformity (DU)


$IU = (Max - Min) / (Max + Min) * 100$, where Max is the maximum and Min is the minimum counts in a pixel

$DU = (Hi - Low) / (Hi + Low) * 100$, where Hi is the highest and Low is the lowest pixel value in a row of 5 pixels moving over the field of view

Matrix size 64x64 or 128x128

UNIFORMITY/DIFFERENT RADIONUCLIDES

Tl 201



Tc 99m

Ga 67

I 131

All 4 images acquired with: Matrix: 256 x 256, counts: 30 Mcounts

Different radionuclides and photon energies

^{99m}Tc



^{201}Tl



^{67}Ga



^{131}I



Uniformity



10% energy window



15% energy window




20% energy window

Uniformity with and without a uniformity correction map

Uniformity



without
uniformity correction map





with
uniformity correction map

- Symmetric energy window — ^{99m}Tc
- Asymmetric energy windows
- Different radionuclides and photon energies
- Uniformity — quantification
- Uniformity — multihead systems
- Corrections (linearity, energy, uniformity)
- Detector crystal
- PM tube and associated electronics
- Collimator
- Artefacts arising from sources/phantoms

Inhomogeneous flood fields can result from:

- Inadequate mixing of Technetium flood phantom
- Computer correction turned off
- Loss of coupling between crystal and PMTs
- Bad PMT voltage adjustment
- Camera pulse height analyzer off-peak for the radionuclide

Uniformity — ^{67}Ga — separate and sum of energy windows



✓ Peak 1 (93 keV), 25% energy window (3.2 million counts).

✓ Peak 2 (184 keV), 20% energy window (1.4 million counts).

✓ Peak 3 (300 keV), 20% energy window (0.4 million counts).

✓ Sum of these three peaks (5 million counts).

First peak




Second peak





Third peak



All peaks




- **Comparison of images acquired with and without a uniformity correction map**






**Comparison of intrinsic
uniformity for ^{201}Tl and ^{67}Ga
on the same scintillation camera
— radionuclide specific
uniformity correction map**

Flood Field



+

**Uniformity
Correction Matrix**




=

**Corrected
Flood Field**



Comparison of images acquired without and with a ^{99m}Tc intrinsic uniformity map correction — defective PM tube





Symmetric energy window — $^{99\text{m}}\text{Tc}$

• A symmetric energy window is one where the photopeak energy is at the centre of the window.


For $^{99\text{m}}\text{Tc}$:

- A 20% energy window set symmetrically over the $^{99\text{m}}\text{Tc}$ photopeak is equivalent to $140 \pm 10\%$ keV or a window spanning 126–154 keV.
- A 15% energy window set symmetrically over the $^{99\text{m}}\text{Tc}$ photopeak is equivalent to $140 \pm 7.5\%$ keV or a window spanning 130–151 keV.
- A 10% energy window set symmetrically over the $^{99\text{m}}\text{Tc}$ photopeak is equivalent to $140 \pm 5\%$ keV or a window spanning 133–147 keV.

Asymmetric window




OFF PEAK MEASUREMENTS



Dogan Bor, Ankara

Intrinsic/Extrinsic Uniformity Quantitative



**Intrinsic Uniformity
IU in UFOV**

**< 3.5
GOOD**

**3.5 – 5
Marginal**

**Above 5
Unacceptable**

**Extrinsic Uniformity
IU in UFOV**

**< 5
GOOD**

**5 – 6
Marginal**

**Above 6
Unacceptable**

**UFOV or CFOV DU of
1.5 -2.0% expected**

System Report



UFOV or CFOV DU of
1.5 -2.0% expected

Full Report of Uniformity Analysis

NAME: 3-04-97 ID: INTR FLD DATE:

UFOV

Integral Uniformity = 2.87%

	Counts	Location
Minimum	7132	(34, 33)
Maximum	7554	(46, 32)

Row Differential Uniformity = 1.69%
Column Differential Uniformity = 1.55%

	Diff.	Location
Max Row	251	(42, 32)
Max Col	230	(46, 28)

CFOV

Integral Uniformity = 2.87%


	Counts	Location
Minimum	7132	(34, 33)
Maximum	7554	(46, 32)

Row Differential Uniformity = 1.69%
Column Differential Uniformity = 1.55%

	Diff.	Location
Max Row	251	(42, 32)
Max Col	230	(46, 28)


Intrinsic/Extrinsic Uniformity

Multi Energy Radionuclides, ^{67}Ga




Each image shows a different non-uniformity. However, the **sum image** is **uniform**. Imaging is acceptable only if all three energy windows are used!

Intrinsic/Extrinsic non Uniformity Artifact



Radius of the ring equals to distance of pixel from center of rotation


Window width



Increased window width will result in an increased number of registered scattered photons and hence a decrease in contrast

Pulse height distribution

The width of the full energy peak (FWHM) is determined by the energy resolution of the gamma camera. There will be an overlap between the scattered photon distribution and the full energy peak, meaning that some scattered photons will be registered.




Energy Resolution

A intrinsic test with 500 micro Ci Tc-99m.

Distance of 5 UFOV again.


Little count acquisition is needed (10 k counts).



Energy resolution =

$(FWHM / \text{Peak energy}) \times 100$

for NaI(Tl), Normal Value is 6-10%







SPATIAL RESOLUTION

- Can be defined in terms of the amount by which a system smears out the image of a very small point source or a very thin line of radioactivity.
- It can be thought of as the distance by which two small point sources must be separated to be distinguished as a separate in the reconstructed image.

- It is usually measured through the use of a transmission phantom (bar phantom)
- It consists of alternating patterns in lead to produce closely spaced areas of differing activity levels, which will allow for the analysis of resolution.
- The better the spatial resolution, the better the ability to detect small abnormalities which will present themselves as different radionuclide concentrations in clinical images.

Spatial resolution


Measured with: Flood source or point source
plus a Bar phantom



Subjective evaluation of the image

- Resolution patterns should be used without a collimator to measure intrinsic performance.
- This can be difficult or impossible along with time consuming with some multihead cameras.
- It can also be useful to assess the resolution with a point or line source.
- The spread of the point or line is an indication of the degree of blurring, or the loss of the resolution in the camera.
- Resolution should be **checked weekly**.

SPATIAL RESOLUTION




Tc-99m or other radionuclide in use

Intrinsic: Collimated line source on the detector

System: Line source at a certain distance

Calculate FWHM of the line spread function



**Four quadrant resolution studies
are usually required weekly.**

**To perform four
quadrant studies you
must rotate the
barphantom 90
degrees so that all
quadrants of the
detector are tested.**





Figure A.5.a. 90° Bar quadrant phantom.



System resolution

Sum of intrinsic resolution and the collimator resolution

Intrinsic resolution depends on:


the positioning of the scintillation events
(detector thickness, number of PM-tubes,
photon energy)

Collimator resolution depends on:


the collimator geometry (size, shape and
length of the holes)

SPATIAL RESOLUTION

Intrinsic resolution



System resolution



SPATIAL RESOLUTION - DISTANCE



Optimal




Large distance

Spatial resolution and linearity

A qualitative index of resolution (performed either intrinsic or extrinsic). Small changes in intrinsic resolution **reduce image contrast**

Test patterns phantoms:



Four-Quadrant Bar Phantom




Orthogonal Hole



Parallel Line Equal Spacing (PLES)

Spatial resolution and linearity – visual





Spatial resolution and linearity

System resolution tends to be a **very stable** parameter

Intrinsic:

Attendant risk to the crystal

Extrinsic:

Moiré patterns (interplay of bar pattern & collimator lead septa)

Intrinsic Spatial resolution and linearity

Nearly the Same as Intrinsic Uniformity test.



1 mCi Tc-99m in 5 UFOV distance from detector.

5-10 million counts with less than 30kcps count rate acquired in a 512x512 matrix size.

Note:

Intrinsic resolution will be 1.75 times the distinguishable smallest bars in the final image.

Intrinsic Spatial resolution



Correct use of these phantoms requires that all images be compared with a reference image (e.g. acceptance test)




Intrinsic Spatial Resolution & linearity (ISRL)

5 mCi Tc-99m

5-10 million counts


30kcps count rate

1024x1024 matrix size




LINEARITY

- The ability to reproduce **a linear activity source as linear** in the image.
- A phantom with either an arrangement of **bars or holes** is usually used.
- The **image produced** should look exactly **like the phantom** that was used.
- **For example:**
 - straight lines should be reproduced as straight lines and holes should be reproduced as holes




LINEARITY




Flood source or point source (Tc-99m)
Bar phantom or orthogonal-hole phantom

1. Subjective evaluation of the image.
2. Calculate absolute (AL) and differential (DL) linearity.



AL: Maximum displacement from ideal grid (mm)
DL: Standard deviation of displacements (mm)

Corrections (linearity, energy, uniformity)




**Illustration of the effect of
no linearity or energy corrections**

LINEARITY

- Linearity should be checked weekly
- **Linear defects** in a flood field are usually the **result of a cracked crystal** and less likely the result of a **collimator defect**.
- Linearity is checked along with spatial resolution with the bar phantoms

- Expressed as **counts/min/MBq** and should be measured for each collimator
- Important to observe with multi-head systems that variations among heads do not exceed 3%



Count Rate Performance

- Performed to **ensure that the time to process** an event is sufficient to maintain **spatial resolution and uniformity in clinical images** acquired at high-count rates

Count Rate Performance


20% Count Rate Loss Test

2 sources of Tc99m
A and B together produce
~20% loss rate



Activity=10% of each other

UFOV lead ring
mask 3 mm thick



Count A, count A and B, count B alone
then reverse - Count B, A and B, then
A and average

Analysis: $A = R_1$
 $A \text{ and } B = R_{12}$
 $B = R_2$

Dead time: $\tau =$


$$\left(\frac{2 R_{12}}{(R_1 + R_2)^2} \right) \ln \left(\frac{R_1 + R_2}{R_{12}} \right)$$

Observed count rate:


$$R_{-20\%} = \frac{0.8}{\tau} \ln \frac{10}{8} = \frac{0.1785}{\tau}$$

Tc99m point source


UFOV lead ring
mask 3mm thick



Rate meter





Remove calibrated
copper absorber
plates




***Intrinsic Spatial Resolution**
measured as before with an
observed count rate of 75,000 cps

***Intrinsic Flood Field Uniformity**
measured as before with an
observed count rate of 75,000 cps

Pixel size



Matrix Size : 256
 Max In profile : 1446
 Distance between points : 200.0 mm
 Distance between points : 141.2 pixels
 X-Pixel size : 1.4165 mm/pixel
 64: 5.6661 128: 2.8330 256: 1.4165 512: 0.7083



Pixel Size Phantom:

Use a small plate of acrylic, with a 1.2 mm x 1.2 mm hole in the center.

Image the phantom in different perpendicular positions with 10 cm distance from center of the detector.

Acquire 50 kcounts within a 1024 x 1024 matrix size, LEGP collimator.

- SPECT uniformity
- Uniformity calibration correction
- Spatial resolution
- Centre of rotation (COR)
- Image alignment in y for multi-head SPECT systems
- Detector head tilt

SPECT performance


- Spatial resolution
- X- and Y-magnification factors and multienergy spatial registration
- Alignment of projection images to axis-of-rotation



- Uniformity
- Camera head tilt

SPECT performance


- ✓ **Reconstruction**
spatial resolution
- ✓ **Reconstruction**
spatial linearity
- ✓ **Reconstruction**
flood field uniformity



Overall system performance

An overall assessment of system performance can be obtained by imaging a suitable tomographic phantom filled with a mixture of water and Tc-99m.

**Jaszczak
SPECT
Phantom**



Overall system performance

The phantom is imaged under ideal conditions:

10 mCi Tc-99m

Minimum radius of rotation,



high resolution collimation,

128 x 128 matrix,

120 views

high total counts (30-50 million)

minimum smoothing recon.



Overall system performance



Thick slices through the **uniform section** of the phantom can be used to evaluate **uniformity** (Ring Artifact).

Thicker slices through the **resolution elements** can be used to determine **tomographic resolution** and **contrast**


(Look for consist transaxial resolution else acquire new COR).

Uniformity (3-5 slices)

Contrast (3 slices)

Resolution (8-12 slices)

SPECT PERFORMANCE



TOMOGRAPHIC UNIFORMITY

Tomographic uniformity is the uniformity of the reconstruction of a slice through a uniform distribution of activity




SPECT phantom with 200-400 MBq Tc99m aligned with the axis of rotation. Acquire 250k counts per angle. Reconstruct the data with a ramp filter

Uniformity


- Nonuniformities that are not apparent in low-count daily uniformity studies can cause significant artifacts in SPECT
- Artifact appears in **transverse images** as **a ring centered** about the AOR
- High-count uniformity images used to **determine pixel correction factors**
 - At least 30 million counts for 64 x 64 images
 - At least 120 million counts for 128 x 128 images

Image of a cylinder filled with a uniform radionuclide solution, showing a **ring artifact due to a nonuniformity**.

Artifact is the dark ring toward the center.



TOMOGRAPHIC RESOLUTION






Method 1: Measurement with the Jaszczak phantom, with and without scatter (phantom filled with water and with no liquid)

Method 2: Measurement with a Point or line source free in air and Point or line source in a SPECT phantom with water

Spatial resolution

- Can be measured by acquiring a SPECT study of a line source (capillary tube filled with a solution of Tc-99m, placed parallel to axis of rotation)
- National Electrical Manufacturers Association (NEMA) specifies a cylindrical plastic water-filled phantom, 22 cm in diameter, containing 3 line sources
- FWHM of the line sources are determined from the reconstructed transverse images (ramp filter)

NEMA phantom for evaluating the spatial resolution of a SPECT scanner




Tomographic spatial resolution in air

Three Point source of ~1 mCi Tc-99m
(less than 2 mm)

Three thin-walled glass capillary tubes
or equivalent shall be used to form
and position the point sources.

These three point sources shall be
made as spherically symmetric as
possible



Tomographic spatial resolution in air

Measurement Procedure:

- The central point source positioned on the **axis of rotation (± 5 mm)**
- Radius of rotation for the circular orbit shall be **100 ± 5 mm**
- Reconstructed matrix has an effective pixel size of **≤ 2.5 mm.**
- **At least 10,000 counts** in each of **128 projections** angle
- Reconstructed Matrix size is **128 x 128**


Calculations and Analysis:


Line profiles across the sources and record the FWHM of the Gaussian fit

Multiply the value with **2.35** to have the FWHM of the tomographic resolution

- NEMA spatial resolution measurements are primarily determined by the collimator used
 - resolution:
 - for peripheral sources**
(7 to 8 mm FWHM collimators)
 - to central resolution**
(9.5 to 12 mm) for LEHR or LEUHR



Gamma camera ring artefacts





CENTER OF ROTATION

MHR/COR alignment/calibration




- Extra High Energy Collimator
- ▲ UHR, HRES, LEAP & Fanbeam Collimator
- All Others

1-5 mCi Tc-99m for each hole
64 projections of 10 seconds
128 x 128 projections
Count rate < 30000 cps

Point source of Tc-99m or Co-57 Make a tomographic acquisition

In x-direction the position will describe a sinus- function. In y-direction a straight line.

Calculate the offset from a fitted sinus and linear function at each angle.



COR

- The *axis of rotation* (AOR) is an imaginary reference line about which the head or heads of a SPECT camera rotate
- If a radioactive line source were placed on the AOR, each projection image would depict a vertical straight line near the center of the image
 - This projection of the AOR into the image is called the *center of rotation* (COR)


COR (cont.)

- Ideally, the COR is aligned with the center, in the x-direction, of each projection image
- Misalignment may be mechanical
 - Camera head may not be exactly centered in the gantry
- Misalignment may also be electronic

CENTER OF ROTATION


Center-of-rotation (COR) misalignment.

Small misalignments cause blurring (*center*), whereas large misalignments cause point sources to appear as “**tiny doughnut**” artifacts (*right*).




Clinical Sample


Quality Control – Gamma camera center of rotation



no offset



offset: 1 pixel



offset: 3 pixel

COR alignment/calibration

Deviations of greater than **0.5 pixel** in a **128 x 128** matrix from the expected position of the source are considered abnormal.

0 pixel



0.5 pixel



1 pixel




2.5 pixel

COR alignment/calibration




The camera heads will rotate in a near **perfect circle** and will remain almost precisely aligned in their opposing positions.....





The predicted or “**electronic**” center of the path of **rotation** **will match** the “mechanical” or actual center of the camera head rotation....

Unfortunately not!



COR alignment/calibration






COR calibration analysis

Center of Rotation
(IAEA TECHDOC 601 8.3.3.)
Mediso Ltd
Serial Number :DH-304031-U0

Head : DH
Isotope ID : Tc-99m
Collimator ID: LEHR
Flood Corr. : INTR
Matrix Size : 64x64x16
View ID : COR_DH
AcqDate : 03.07.2005
AcqStartTime : 14:08:47
Total Time : 10.0 secs
Sum : 60416
Max.pix value: 10229

	[mm]	[pixel]
X offset :	-0.0423	[-0.0051]
dX max :	0.2281	[0.0276]
dX min :	-0.1876	[-0.0227]
dY max :	0.4259	[0.0516]
dY min :	-0.3037	[-0.0368]
COR :	0.4259	[0.0516]




COR calibration analysis

TL: Angular x position
sinogram of the
calibration point source.

BL: Angular position
difference between the
sinogram and a fitted
sine wave.

TR: Angular y position of
the point.



BR: Angular variation of
the COR value.



Camera head tilt

- Camera head or heads must **be exactly parallel to the AOR**
 - If not, **loss of spatial resolution and contrast**
 - Loss will be less toward the center of the image and greatest toward the edge of the image
- Can assess **using a point source** in camera's FOV, **centered in the axial (y) direction, but near the edge in the transverse (x) direction**
 - If there is head tilt, position of point source will vary in y-direction from image to image

Tilted detector



COR calibration

- **Misalignment** may be corrected by shifting each image in the x-direction by the proper number of pixels prior to filtered backprojection
- If **COR misalignment** varies with camera head angle, it can only be corrected if computer permits angle-by-angle corrections
- Separate assessments of **COR correction** must be made for different collimators

Conclusion

Test	Purpose	Frequency
Physical inspection	Collimator mounting and damage	Daily
Collimator touch pad and gantry emergency stops	Unexpected collision with the patient or an obstacle during motion	Daily, and after collimator change
Energy window setting	To check and center the energy window on the photopeak	Daily, for any radionuclide
Background count rate	Detect radioactive contamination	Daily
Intrinsic/extrinsic uniformity and sensitivity- visual	Test response to uniform flux of photons, record cps/MBq and monitor sensitivity	Daily (low count)
Intrinsic/extrinsic uniformity and sensitivity- quantitative	To monitor the trend in uniformity with quantitative indices, check the sensitivity	Weekly/monthly (high count)
Spatial resolution and linearity – visual	To detect distortion of spatial resolution and linearity	Six-monthly
Multiple window spatial registration (MWSR)	To test that all images superimpose in an additive or subtractive mode	Six-monthly yearly

Conclusion

Test	Purpose	Frequency
Detector head tilt	To adjust the alignment of detector head tilt in the Y-axis	Before use
COR alignment/calibration	To check that the mechanical and electronic CORs are aligned	Weekly/As required
Overall system performance	Tomographic uniformity and contrast resolution	Six-monthly
Tomographic spatial resolution in air	To check tomographic spatial resolution of the system in air, with no scatter	Six-monthly
Pixel size	To determine absolute pixel size for quantitation, fusion and AC	Six-monthly

**Thank you for
your Attention!**



SPECT/CT Artifacts & Solutions			
Artifact Type	Cause	Appearance	Solution
COR Offset	Gantry-detector misalignment	Blurring, double images	Recalibrate COR weekly (<0.5 mm error)
Bubble Artifact	Air trapped in PMT-crystal gap	Circular cold spots in flood images	Reseal or replace detector module
Detector Hydration	Moisture in NaI(Tl) crystal	Yellowed crystal, patchy uniformity	Replace crystal; control humidity
Ramp Filter Artifact	Over-filtering in FBP	Star-like noise streaks	Use Butterworth filter or switch to OSEM
Reconstruction Error	Wrong OSEM iterations/ μ -map	Checkerboard pattern, false defects	Optimize iterations/subsets; verify CT-SPECT alignment
Gantry Misalignment	Mechanical tilt or loose parts	Asymmetric blurring	Relevel detector; mechanical service
Donut Artifact	Severe non-uniformity/COR error	Concentric rings in SPECT	Recalibrate uniformity and COR
Attenuation Artifact	Misaligned CT μ -map	False cold spots near bones/metal	Verify CT-SPECT registration; use MAR algorithms
Scatter Artifact	Wide energy window	Halo around high-uptake organs	Narrow window ($\pm 10\%$ for Tc-99m); apply scatter correction
Edge Packing	Aggressive edge enhancement	Artificial bright edges	Reduce high-pass filtering
Photopeak Offset	Energy calibration drift	Global non-uniformity	Daily energy peak recalibration
Fusion Misregistration	Patient motion/gantry error	Misaligned anatomy (e.g., heart vs. ribs)	Immobilize patient; manual re-fusion
HU-to- μ Map Error	Uncalibrated CT attenuation	Artificial cold/hot spots	Calibrate CT-SPECT μ -conversion monthly
Non-Uniformity (SUV Errors)	Detector calibration skipped	Inaccurate quantification	Daily uniformity flood correction









Spirit DH-V v7.01

Uniformity test
(NEMA NU 1-1994.2.4)
Rajai Heart Center
Serial Number :DH-304031-V0

Head : 2
Isotope ID : Tc-99m
Collimator ID: INTR
Flood Corr. : INTR
Matrix Size : 256x256x16
View ID : FLOOD_TEST_IN_H2
AcqDate : 02.07.2009
AcqStartTime : 18:10:34
Total Time : 580.6 secs
Sum : 16000004
Max.pix value: 1408

Flood field uniformity

	UFOV	CFOV
Diff. unif.:	17.6 %	17.6 %
Int. unif.:	30.2 %	28.8 %
Average :	5438	5383
Maximum :	6747	6543
Minimum :	3616	3616



ESC: Go Back

F3: Proceed

PgDn/PgUp: Palette




Spirit DH-U v7.01

Uniformity test
(NEMA NU 1-1994.2.4)
Rajai Heart Center
Serial Number :DH-304031-U0

Head : 2
Isotope ID : Tc-99m
Collimator ID: INTR
Flood Corr. : INTR
Matrix Size : 256x256x16
View ID : FLOOD_TEST_IN_H2
AcqDate : 01.31.2009
AcqStartTime : 20:37:05
Total Time : 299.0 secs
Sum : 16000050
Max.pix value: 807

Flood field uniformity

	UFOV	CFOV
Diff. unif.:	1.7 %	1.7 %
Int. unif.:	2.1 %	2.0 %
Average	5511	5508
Maximum	5616	5616
Minimum	5390	5401



ESC: Go Back F3: Proceed PgDn/PgUp: Palette




Spirit DH-U v7.01

Uniformity test
(NEMA NU 1-1994.2.4)
Rajai Heart Center
Serial Number : DH-304031-U0

Head : 2
Isotope ID : Tc-99m
Collimator ID: INTR
Flood Corr. : INTR
Matrix Size : 256x256x16
View ID : FLOOD_TEST_IN_H2
AcqDate : 05.02.2009
AcqStartTime : 19:31:58
Total Time : 217.9 secs
Sum : 10750383
Max.pix value: 308

Flood field uniformity

	UFOV	CFOV
Diff. unif.:	6.1 %	4.8 %
Int. unif.:	7.1 %	5.0 %
Average :	3747	3765
Maximum :	3957	3957
Minimum :	3430	3582




ESC: Go Back F3: Proceed PgDn/PgUp: Palette

Spirit DH-U v7.01

Uniformity test
(NEMA NU 1-1994.2.4)
Rajai Heart Center
Serial Number :DH-304031-U0


Head : 1
Isotope ID : Tc-99m
Collimator ID: INTR
Flood Corr. : INTR
Matrix Size : 256x256x16
View ID : FLOOD_TEST_IN_H1
AcqDate : 05.02.2009
AcqStartTime : 19:25:18
Total Time : 171.8 secs
Sum : 2150415
Max.pix value: 261




ESC: Go Back


F3: Proceed

PgDn/PgUp: Palette










Non Uniform





Uniform




Extrinsic
Uniformity



128×128







512×512




64×64

Unusual Artifacts




Correction Tables


No
corrections




Energy
And
Linearity



Energy
only




Energy,
Linearity,
Uniformity
(all corrections)




^{99m}Tc Intrinsic Flood Images

Correction Tables


Energy
only




Energy and
Uniformity,
No Linearity
Correction



Energy
And
Linearity
Only




All Corrections
On



^{99m}Tc Intrinsic


Intrinsic Bars – Linearity Correction Off



Flood Images – Off Peak

CLASS IMAGES 4-23


23Apr2007



HD1 HIGH BY 7

CLASS IMAGES 4-23


23Apr2007



HD2 LOW BY 7

CLASS IMAGES 4-23


23Apr2007



HD1 HI BY 14

CLASS IMAGES 4-23


23Apr2007



HD2 LO BY 14

Low to High Count Rate Intrinsic Floods


19 kcps



DET 2 TC LO

04/16/09 17:05:49


79 kcps



DET 2 TC HIGH2

04/16/09 17:22:48

109 kcps
(too high)




DET 2 TC HIGH3

04/16/09 17:26:47

Tube-like Artifacts visible in Parathyroid study

NOT FOR DIAGNOSTIC PURPOSES



DET 1

04/30/09 10:20:22

Cause: ^{99m}Tc Aerosol from previous patient pulled into camera head by fans






Image made with no other source other than aerosol pulled into camera heads. Tc present behind the crystal. Patient breathed aerosol close to camera heads and leakage occurred around breathing apparatus.

Daily QC – Water filled sheet source with ^{99m}Tc MAA added


Morning QC looked bad

QC repeated with same source later in afternoon




Note “clumping” effect caused by using MAA.
 ^{99m}Tc MAA (normally used for lung perfusion imaging) was used by mistake. Should use ^{99m}Tc pertechnetate.
Sheet source had to be emptied and rinsed

Bar pattern – intrinsic vs. extrinsic




Intrinsic – better resolution
than extrinsic




Extrinsic

Intrinsic Bar Pattern Tc and Thallium




^{99m}Tc




^{201}Tl

Intrinsic Thallium bar pattern, One Peak at a time




Lower energy peak only, 69 keV



Upper energy peak only, 167 keV –
Better resolution at higher energy

Bar Pattern with wrong collimator Medium Energy Collimator




Bar pattern – slight nonlinearity in corner








Ring Artifacts




←
Ring artifacts
visible


Ring Artifacts



Ring Artifacts



Severe Ring Artifacts




SPECT Ring Artifacts

- Caused by non-uniformities such as:
 - Visible non-uniformities in flood image due to camera being off peak, PMT gain imbalance, or need for new correction tables
 - Shift in photopeak as camera head rotates
 - Collimator defect or damage (not visible in intrinsic flood image)
- Even small non-uniformities can cause ring artifacts

**Thank you for
your Attention!**



NOISE



Count density

WHOLE BODY

- Scan speed
- Image size
- Multiple scan paths
- Detector uniformity
- Energy window Radioactive contamination

Multiple Window Spatial Registration

Multi-peak imaging with isotopes such as **Ga-67** or **In-111** assumes images acquired at **different energies** are directly **superimposable**.

Unfortunately not!

Magnitude of the positioning signals generated by the pulse arithmetic circuitry changes with gamma ray energy!



SO:


These signals are normalized to some reference energy to insure that the position of an event is independent of gamma ray energy.

This correction is achieved in term of **MWSR**!

Multiple Window Spatial Registration

Image a set of 5 collimated point sources of Ga-67 (or Ba-133) at their different photopeaks acquired at 1024 x 1024 matrix size.

Subtract the 93 keV from the 296 keV image, is a **simple visual** indication of the co-registration of the two images.



The worst difference between location of 2 different energies is **MWSR!** In modern gamma cameras, the use of energy and linearity correction maps has significantly improved the MWSR.